



# LUNG CENTER OF THE PHILIPPINES

Quezon Avenue Extension, Quezon City, Philippines 1100

ISO 9001:2008 CERTIFIED

Date : May 03, 2018

TO : ALL PROSPECTIVE BIDDERS  
PROCUREMENT FOR 1 YEAR SUPPLY OF VARIOUS REAGENTS FOR CLINICAL CHEMISTRY AND MICROBIOLOGY SECTION USE

FROM : BIDS AND AWARDS COMMITTEE


SUBJECT : SUPPLEMENTAL BID BULLETIN NO. 1  
"MODIFICATIONS ON THE SPECIFICATIONS FOR ITEMS NO 1 AND 3 OF CLINICAL CHEMISTRY AND CLARIFICATION ON THE TERMS OF REFERENCE

This Bid Bulletin No. 1 is issued to inform All Prospective Bidders of the modifications on the specifications and Terms of Reference for Clinical Chemistry.

1. Wet Clinical Chemistry Analyzer
  - Has the ability to manually or automatically remove empty reagents when their on-board life has expired
2. Hemoglobin A1c Analyzer
  - Touch screen operation/Built in or incorporated key pad/and or PC operated thru windows
3. Provides free lifetime preventive maintenance/machine calibration with provision of Certificate thereafter

Strictly follows schedule PM/Calibration with free parts and labor especially during machine breakdown

For information and guidance of all concerned.

  
**MS. ELVIRA N. BAURA, RN**  
Chairman, LCP BAC